


REVISIONS			
REV	DESCRIPTION	DATE	APPROVED

- 1 ISO: NW50 KF Flange
 2 4-40 UNC x .160 deep 4 places



Accu-Glass Products Inc.
 25047 Anza Drive
 Valencia, CA 91355
 TEL 818-365-4215
 FAX 818-365-7074
 accuglassproducts.com

Notes :

*** UNCONTROLLED WEB COPY ***

Material
 304 SS

Tolerances unless
 otherwise specified
 Decimal Angular
 .xx= +/- .005 +/- 1°
 .xxx= +/- .002

Title
 15D-4TC2P-K50
 HV TC-Power Feedthrough

Drawn By	M.Wilson	DWG NO.	100801	Date	7-31-97
Approved		PDFE		Cage Code : 4HCQ8	