




Notes

- 1. Half-Nipple 6.0 with ISO NW160 LF Flange
- 2. Protect flange O-ring seal surface from damage
- 3. Leak test to 5x10E-10 Standard cc/sec He

Rev.	Description	Date	Made By	Approval
A	Initial Release			



**Accu-Glass Products Inc.**  
 25047 Anza Drive  
 Valencia, CA 91355  
 T 818-365-4215  
 F 818-365-7074  
 www.accuglassproducts.com

**UNCONTROLLED WEB DRAWING**

Proprietary and Confidential  
 The information contained in this drawing is the sole property of Accu-Glass, Products Inc. Any reproduction, in part or as whole, without the written permission of Accu-Glass Products, Inc., is strictly prohibited.

Material Stainless Steel, 304	
Tolerances, unless otherwise specified	
.X ± .030	Angles ± 0.5°
.XX ± .010	Finish 63Ra
.XXX ± .005	

Title Nipple, Half / 6.0 ISO NW160 LF Flange				
Drawn by Date	Ligeti 10/1/2018	No. <b>210604</b>	Rev. <b>A</b>	
Approved			Sheet	

A

B

C

D